

MyoFetch2019 - Task #14519

Task # 14470 (New): Variable Monitoring
Task # 14504 (In Progress): EMG processing circuit

Filter Design

07/04/2019 23:12 - Duarte Pereira

Status:	Resolved	Start date:	01/04/2019
Priority:	High	Due date:	13/05/2019
Assignee:	Duarte Pereira	% Done:	95%
Category:		Estimated time:	0.00 hour
Target version:		Spent time:	0.00 hour
Description Designing the Low Pass Filter (cut off frequency of 20Hz) and the High Pass Filter (cut off frequency of 709Hz), as sugested in the following article: https://pdfs.semanticscholar.org/c348/f06416aa03154c40bb82d5b26cef40e5840d.pdf - last accessed 07/04/2019			
Subtasks:			
Task # 14525: Filter Testing			Resolved
Task # 14580: Obtaining a readable EMG signal			Resolved

History

#1 - 07/04/2019 23:19 - Duarte Pereira

- % Done changed from 0 to 100

#2 - 08/04/2019 14:37 - Duarte Pereira

- File deleted (Filters.zip)

#3 - 08/04/2019 14:38 - Duarte Pereira

- File Filter_Design_Testing.zip added

Files

Filter_Design_Testing.zip	5.83 MB	08/04/2019	Duarte Pereira
---------------------------	---------	------------	----------------